

INDUSTRIAL COMPONENT DATA

078214 LEGRAND TABLE 36 ADDRESSES

Brand: LEGRAND | Manufacturer Part Number: 078214



Manufacturer	LEGRAND
MPN / Reference	078214
Catalog number	078214
Category	Accessories/spare parts for controls
Minimum order	1 Pcs
Document ID	EO-7380EB6856
Generated	2026-07-29

Product Overview

TABLE 36 ADDRESSES

Table control unit - up to 6 display units Cat. No 0 766 60 - 36 modules

Mosaic nurse call - table desk for 6 indicators max

STRUCTURED PRODUCT DATA

Technical Specifications

No structured technical attributes are currently available. Submit an RFQ for technical review.

Classification & Catalog Path

Catalog path	LEGRAND > SOLUTIONS FOR SANITARY FACILITIES
--------------	---

Additional Product Information

Pack (number of units)1

Volume (dm3)11.00

Weight (g)1 321.00

Product characteristicsDisplay units and management

Table control unit

Takes up to 6 display units Cat.No 0 766 60

Horizontal or vertical mounting

Dim.: 310 x 295 x 75 mm

36 modules

General characteristicsMosaic™ Programme nurse call units - for day hospitals, dialysis centres...Designed for nurse call equipment in day hospitals, dialysis and transfusion centres, scanners, etc.

Functions:from the room: calls, emergency calls, nurse present indicator and call cancelling

from the nurses' station: patient reassurance

Possibility of adding a paging system

ETIM for 078214 PLC's (EC002584)

Feature Name

Feature Value

Type of mechanical accessory/spare part (EF005994)

Other

Type of electrical accessory/spare part (EF005722)

Other

REQUEST PRICE, AVAILABILITY & TECHNICAL REVIEW

Send LEGRAND / 078214 with the required quantity to info@electricaonline.com or use electricaonline.com/request-quote.

GLOBAL OFFICES

Germany: Lise-Meitner-Straße 13, 89081 Ulm | USA: 16225 Park Ten Pl #500, Houston, TX 77084 | UAE: HDS Business Centre Tower, Cluster M, 33rd Floor, JLT, Dubai

Catalog-generated document. Product names, marks and specifications belong to their respective owners. Verify safety-critical, compatibility and certification requirements against the manufacturer documentation before purchase or installation.